

SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 01/08/2025


OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:12
INTERNAL STANDARD	VERIFIED	07:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:13
BLANK TEST	0.00	07:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:14
BLANK TEST	0.00	07:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:15
BLANK TEST	0.00	07:16

1/8/25

MC 20



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 01/08/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:12
INTERNAL STANDARD	VERIFIED	07:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:13
BLANK TEST	0.00	07:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:14
BLANK TEST	0.00	07:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:15
BLANK TEST	0.00	07:16

1/8/25
MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 01/16/2025




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:46
INTERNAL STANDARD	VERIFIED	00:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:47
BLANK TEST	0.00	00:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:48
BLANK TEST	0.00	00:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:49
BLANK TEST	0.00	00:50

1/16/25

MC 20



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/27/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:50
INTERNAL STANDARD	VERIFIED	00:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:51
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:52
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:53
BLANK TEST	0.00	00:54

1/27/25
MC520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 01/31/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:21
INTERNAL STANDARD	VERIFIED	07:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:21
BLANK TEST	0.00	07:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST	0.00	07:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:23
BLANK TEST	0.00	07:24

1/31/25
MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 02/06/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:32
INTERNAL STANDARD	VERIFIED	00:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:32
BLANK TEST	0.00	00:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:33
BLANK TEST	0.00	00:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:34
BLANK TEST	0.00	00:35

2/6/25
MC 520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 02/21/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:06
INTERNAL STANDARD	VERIFIED	00:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:06
BLANK TEST	0.00	00:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:08
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:09
BLANK TEST	0.00	00:10

2/21/25
Mc520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 03/01/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:25
INTERNAL STANDARD	VERIFIED	01:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:26
BLANK TEST	0.00	01:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:27
BLANK TEST	0.00	01:28

3/1/25
McF20



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 03/08/2025

OPERATOR NAME:
JOHN M GENITO
Permit #: 53620
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:15
INTERNAL STANDARD	VERIFIED	07:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:15
BLANK TEST	0.00	07:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:16
BLANK TEST	0.00	07:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:17
BLANK TEST	0.00	07:18

JS 538
3/8/2025



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 03/16/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:12
INTERNAL STANDARD	VERIFIED	01:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:12
BLANK TEST	0.00	01:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:13
BLANK TEST	0.00	01:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:14
BLANK TEST	0.00	01:15

3/16/25
MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 03/16/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:42
INTERNAL STANDARD	VERIFIED	07:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:42
BLANK TEST	0.00	07:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:43
BLANK TEST	0.00	07:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:44
BLANK TEST	0.00	07:45

MC
3/16/25



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 03/29/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:06
INTERNAL STANDARD	VERIFIED	00:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:06
BLANK TEST	0.00	00:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:07
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:08
BLANK TEST	0.00	00:09

3/29/25
MC



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 03/31/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:47
INTERNAL STANDARD	VERIFIED	23:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:47
BLANK TEST	0.00	23:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:48
BLANK TEST	0.00	23:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:49
BLANK TEST	0.00	23:50

Handwritten:
AC520
3/31/25



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 04/08/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:01
INTERNAL STANDARD	VERIFIED	07:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:01
BLANK TEST	0.00	07:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:02
BLANK TEST	0.00	07:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:03
BLANK TEST	0.00	07:04

4/8/25
MCB



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 04/16/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:14
INTERNAL STANDARD	VERIFIED	23:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:14
BLANK TEST	0.00	23:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:16
BLANK TEST	0.00	23:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:17
BLANK TEST	0.00	23:18

4/16/25
MC 20



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 04/27/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:32
INTERNAL STANDARD	VERIFIED	06:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:32
BLANK TEST	0.00	06:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:33
BLANK TEST	0.00	06:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:34
BLANK TEST	0.00	06:35

4/27/25
MC 520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 05/01/2025




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:09
INTERNAL STANDARD	VERIFIED	00:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:09
BLANK TEST	0.00	00:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:11
BLANK TEST	0.00	00:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:12
BLANK TEST	0.00	00:12

5/1/25

MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 05/09/2025

OPERATOR NAME:
JOHN M GENITO
Permit #: 53620
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:42
INTERNAL STANDARD	VERIFIED	22:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:43
BLANK TEST	0.00	22:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:44
BLANK TEST	0.00	22:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:45
BLANK TEST	0.00	22:46

J6538
5/9/25



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 05/18/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:33
INTERNAL STANDARD	VERIFIED	06:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:33
BLANK TEST	0.00	06:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:34
BLANK TEST	0.00	06:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:35
BLANK TEST	0.00	06:36

5/18/25
MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 05/26/2025

OPERATOR NAME:
JOHN M GENITO
Permit #: 53620
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:45
INTERNAL STANDARD	VERIFIED	05:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:46
BLANK TEST	0.00	05:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:47
BLANK TEST	0.00	05:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:48
BLANK TEST	0.00	05:49

5/26/25
JMG 538



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 05/30/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:22
INTERNAL STANDARD	VERIFIED	01:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:22
BLANK TEST	0.00	01:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:23
BLANK TEST	0.00	01:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:24
BLANK TEST	0.00	01:25

5/30/25
MC520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 06/05/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:45
INTERNAL STANDARD	VERIFIED	00:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:45
BLANK TEST	0.00	00:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:46
BLANK TEST	0.00	00:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:47
BLANK TEST	0.00	00:48

6/5/25
MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 06/15/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:48
INTERNAL STANDARD	VERIFIED	05:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:49
BLANK TEST	0.00	05:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:50
BLANK TEST	0.00	05:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:51
BLANK TEST	0.00	05:52

n/c 6/15/25



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 06/22/2025

OPERATOR NAME:
JAMES TESSEYMAN
Permit #: 47783
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:55
INTERNAL STANDARD	VERIFIED	18:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:56
BLANK TEST	0.00	18:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:57
BLANK TEST	0.00	18:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	18:58
BLANK TEST	0.00	18:59

06/22/2025
JTS/92



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 06/30/2025

OPERATOR NAME:
MICHAEL P MCRORY
Permit #: 53626
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:26
INTERNAL STANDARD	VERIFIED	22:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:26
BLANK TEST	0.00	22:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:27
BLANK TEST	0.00	22:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:29
BLANK TEST	0.00	22:29

MP MCRORY #535
6/30/2025



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 06/30/2025

OPERATOR NAME:
MICHAEL P MCRORY
Permit #: 53626
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:32
INTERNAL STANDARD	VERIFIED	22:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:33
BLANK TEST	0.00	22:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:34
BLANK TEST	0.00	22:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:35
BLANK TEST	0.00	22:36

6/30/2025
10/10/25 #535



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 07/12/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:36
INTERNAL STANDARD	VERIFIED	05:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:36
BLANK TEST	0.00	05:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:37
BLANK TEST	0.00	05:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:38
BLANK TEST	0.00	05:39

7/12/25
CS20



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306



Date: 07/21/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:44
INTERNAL STANDARD	VERIFIED	01:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:44
BLANK TEST	0.00	01:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:46
BLANK TEST	0.00	01:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:48

*7/21/25
MC/20*



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 07/30/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:13
INTERNAL STANDARD	VERIFIED	06:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:13
BLANK TEST	0.00	06:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:15
BLANK TEST	0.00	06:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:16
BLANK TEST	0.00	06:16

7/30/25
MC 20



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 08/09/2025

OPERATOR NAME:
JOSEPH GALVIN
Permit #: 45745
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:30
INTERNAL STANDARD	VERIFIED	11:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:31
BLANK TEST	0.00	11:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:32
BLANK TEST	0.00	11:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:33
BLANK TEST	0.00	11:34

8/9/25
SG 493



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 08/19/2025


OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:10
INTERNAL STANDARD	VERIFIED	22:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:10
BLANK TEST	0.00	22:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:11
BLANK TEST	0.00	22:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:12
BLANK TEST	0.00	22:13

8/19/25

Mc 520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 08/28/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:13
INTERNAL STANDARD	VERIFIED	07:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:14
BLANK TEST	0.00	07:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:15
BLANK TEST	0.00	07:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:16
BLANK TEST	0.00	07:17

8/28/25
Mc 520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 08/29/2025




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:06
INTERNAL STANDARD	VERIFIED	19:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:07
BLANK TEST	0.00	19:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:08
BLANK TEST	0.00	19:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:09
BLANK TEST	0.00	19:10

8/29/25

Mc 520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 09/17/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:45
INTERNAL STANDARD	VERIFIED	05:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:46
BLANK TEST	0.00	05:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:47
BLANK TEST	0.00	05:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:48
BLANK TEST	0.00	05:49

9/17/25
MC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 09/24/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:20
INTERNAL STANDARD	VERIFIED	22:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:20
BLANK TEST	0.00	22:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:21
BLANK TEST	0.00	22:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:22
BLANK TEST	0.00	22:23

9/24/25
Mc520



SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 10/01/2025




OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:54
INTERNAL STANDARD	VERIFIED	00:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:54
BLANK TEST	0.00	00:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:55
BLANK TEST	0.00	00:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:57
BLANK TEST	0.00	00:57

10/1/25

UC520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 10/11/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:37
INTERNAL STANDARD	VERIFIED	23:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:38
BLANK TEST	0.00	23:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:39
BLANK TEST	0.00	23:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:40
BLANK TEST	0.00	23:41

10/11/25
MC520



SUPERVISOR TEST




RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 10/27/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:01
INTERNAL STANDARD	VERIFIED	01:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:01
BLANK TEST	0.00	01:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:02
BLANK TEST	0.00	01:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:03
BLANK TEST	0.00	01:04



10/27/25
MCS20

SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 11/02/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:21
INTERNAL STANDARD	VERIFIED	01:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:21
BLANK TEST	0.00	01:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:23
BLANK TEST	0.00	01:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:24
BLANK TEST	0.00	01:25

11/2/25
MCF20



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 11/10/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:07
INTERNAL STANDARD	VERIFIED	01:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:08
BLANK TEST	0.00	01:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:09
BLANK TEST	0.00	01:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:10
BLANK TEST	0.00	01:11

McS20
11/10/25



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 11/17/2025

OPERATOR NAME:
JAMES TESSEYMAN
Permit #: 47783
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:11
INTERNAL STANDARD	VERIFIED	20:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:11
BLANK TEST	0.00	20:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:12
BLANK TEST	0.00	20:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	20:13
BLANK TEST	0.00	20:14

11/17/2025
J492



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 11/17/2025

OPERATOR NAME:
JAMES TESSEYMAN
Permit #: 47783
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:44
INTERNAL STANDARD	VERIFIED	18:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD INTERFERENCE		18:45

Interference Detected



11/17/2025

JT498

SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 11/17/2025

OPERATOR NAME:
JAMES TESSEYMAN
Permit #: 47783
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:38
INTERNAL STANDARD	VERIFIED	18:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	INTERFERENCE	18:39

Interference Detected



11/17/2025
JTE

SOLUTION CHANGE

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306


Date: 12/12/2025

OPERATOR NAME:
JOSEPH T GALVIN
Permit #: 45745
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:51
INTERNAL STANDARD	VERIFIED	19:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:51
BLANK TEST	0.00	19:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:53
BLANK TEST	0.00	19:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:54
BLANK TEST	0.00	19:54

12/12
JB493



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306




Date: 12/21/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:19
INTERNAL STANDARD	VERIFIED	08:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:20
BLANK TEST	0.00	08:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:21
BLANK TEST	0.00	08:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:22
BLANK TEST	0.00	08:23

12/21/25
MC 520



SUPERVISOR TEST

RAMAPO TOWN
POLICE DEPT
DATAMASTER dmt: 116306

Date: 12/31/2025

OPERATOR NAME:
MICHAEL S CARBONE
Permit #: 50609
Operator Agency/Dept: RAMAPO TOWN PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:18
INTERNAL STANDARD	VERIFIED	01:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:21
BLANK TEST	0.00	01:22

12/31/25
MC 520

